

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

05 2011

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>180194</u>
Name	<u>Jaime Smith</u>	ID# <u>6762</u> Date <u>6/29/11</u>
A	Agency <u>Denali National Park</u>	Phone # <u>907-683-9525</u>
Instrument Location <u>Denali National Park - Headquarters</u>		
B		
Alco S/N	<u>X301591</u>	Target Value <u>0.080</u> High Pressure <u>1200 PSI</u>
Alco Test Values	<u>.077</u> 1 <sup>st</sup> Alco	<u>.077</u> 2 <sup>nd</sup> Alco
Signature	<u>Jaime Smith</u>	
(OVER)		

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 22 day of July, 2010.

Nikki Roth  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office

(Notary Seal)



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130194

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130194

JUNE 29, 2011

OPERATOR'S NAME:

SMITH/JACQUEL

INSTRUMENT NUMBER: 6452

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/ID :

X

O.L. #: X

DEPT/AGENCY: DMLB

CASE/REPORT:

TEST TYPE: V

ALCO TEST VALUE: 0.052

ALCO S/N: 1301531

### --- BREATH ANALYSIS ---

ALCO ADJUSTED FOR 27.59 in

ALCO CHECK .073 09:23

BLANK TEST .000 09:04

INTERNAL STANDARD VERIFIED 09:05

ALCO @ 27.59 in .077 09:05

BLANK TEST .000 09:06

SUBJECT SAMPLE .000 09:06

BLANK TEST .000 09:07

ALCO @ 27.59 in .077 09:07

BLANK TEST .000 09:08

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130194

JUNE 29, 2011

TIME 08:09

### --- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 33c

BREATH TUBE: 41c

BAROMETER: 27.59 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

ALCO TEST VALUE: 0.052  
ALCO S/N: 1301531  
SUBJECT'S LAST NAME: JACQUEL  
SUBJECT'S FIRST NAME/ID: X

